



Health-Chem Diagnostics, LLC

3341 S.W 15th ST. POMPANO BEACH, FL 33069 – USA (954) 979-3845 – Fax: (954) 979-7997 www.healthchemdiagnostics.com

BREATHCHECK™ BREATH ALCOHOL DETECTOR

INSTRUCTIONS

Wait 15 minutes after last alcoholic drink (alcohol takes about 15 minutes to have an effect on your system), OR drink a glass of water before taking the test.

	<p>1. Squeeze middle of the outer plastic tube between thumb and forefinger to break the inner glass ampoule containing yellow crystals. Squeeze only once. Do not crush or bend tube. Use tester immediately.</p>
<p>2. 12 seconds</p> 	<p>2. Take a deep breath and blow in one continuous breath for 12 seconds through the end of the tube designated by arrow. BLOW VERY HARD – EXHALE THROUGH THE TUBE. DO NOT INHALE. Shake tester to distribute crystals evenly, wait two (2) minutes. Identify color change of majority of crystals.</p>
<p>After 2-3 minutes, compare the color of the majority of the crystals to the graduated scale on the label of the tester. Match the color of the granules against the color scale on the label for the resultant concentration.</p>	

INTERPRETING YOUR RESULTS

Note: Use the Color Chart on the BreathCheck device itself for greatest accuracy

0.0%	NEGATIVE RESULT
0.04%	<i>If 0.04% comes up, since alcohol varies with time, this may be considered positive.</i> NOT RECOMMENDED TO DRIVE
0.08%	High alcohol level DO NOT DRIVE!

NOTE: Accuracy of the test results may not be reliable if the test is not conducted according to the instructions. **REFRAIN** from driving after drinking – **DO NOT DRIVE AT ALL** if any crystals change color.

PRODUCT DESCRIPTION

BreathCheck - Breath Alcohol Detector is a disposable breath alcohol indicator designed for one-time use. The BreathCheck - Breath Alcohol Detector contains indicator chemistry that will undergo a color change in the presence of alcohol contained in the breath of the subject. This product provides a reliable indication of alcohol present in the exhaled breath of the test subject when the instructions for use of the BreathCheck - Breath Alcohol Detector are rigidly followed.

The manufacturers, suppliers, agents, distributors and retailers make no warranty expressed or implied as to the ability of this device to determine or detect intoxication of the subject or to accurately indicate the subject's blood alcohol level. Decisions and/or actions based on the use of this product by any person shall be at such persons own risks. The manufacturers, suppliers, agents, distributors and retailers assume no responsibility for consequences of subjects who test negative using this device, but who later show that they are under the influence of, or their judgment has been impaired by alcohol.

WARNING: This product should be used only as a screening device and is only an indication of the possible presence of alcohol in the blood of the test subject. Correlation between breath alcohol content and blood alcohol content depends on many variables, including altitude. The exact concentration of alcohol in the blood of the test subject cannot be accurately determined by using this device. This device is not intended to legally determine blood alcohol presence or level. No inference of intoxication is to be made from a positive indication. BreathCheck - Breath Alcohol Detector is guaranteed to be free from manufacturer's defects. This warranty is expressly made in lieu of any and all other warranties expressed or implied including the warranties of merchantability and fitness for a particular purpose or use. There are no warranties expressed beyond the description of the product contained on this package. The warrantor expressly disclaims liability for incidental, special or consequential damages of any nature.

Use immediately after breaking glass vessel. Do not use if glass vessel containing crystals is ruptured prematurely or if crystals are not yellow.

Keep out of reach of children. Do not immerse in liquid.

Do not ingest. If ingested, induce vomiting and contact your physician.

Note: Should not be used in Sodium Vapor Light conditions which could alter the appearance of the color.

Manufactured in Pompano Beach, Florida - USA by:
HEALTH-CHEM DIAGNOSTICS LLC

www.healthchemdiagnostics.com
Certified ISO CMDCAS 13485:2003



FM77504 - Quality Award
FDA Registration No.1048532