

**CERTIFICATE OF CONFORMANCE
EUROPEAN COMMUNITY
COUNCIL DIRECTIVE 93/42/EEC
(EMC Aspects Only)**

Date of Issue: June 2009

Issued By: Retlif Testing Laboratories
795 Marconi Avenue
Ronkonkoma, NY 11779

Issued To: Weiss Instruments, Inc.
905 Waverly Avenue
Holtsville, NY 11742

Reference: Retlif Report Number R-12991

Retlif Testing Laboratories hereby acknowledges that compliance testing in accordance with the below listed standards was performed on a representative sample of the equipment listed below. Retlif Testing Laboratories further acknowledges that the test sample listed below was found to be in compliance with these standards.

This certificate is hereby issued to the above named grantee and is valid only for the equipment identified below.

Manufacturer: Weiss Instruments, Inc.
905 Waverly Avenue
Holtsville, NY 11742

Equipment Tested: TD Glucose Meter

Serial Numbers: B037, B039

Brand Name: Health-Chem Electronics

Product Type: Medical Equipment

Note(s):

- 1) See attached report R-12991 for details and/or conditions pertaining to this certificate.
- 2) Conforms to the emissions requirements of IEC 60601-1-2:2007
CISPR 11 Edition 4: 2003, A1: 2004, A2: 2006 Radiated Emissions, Group 1, Class B
- 3) Conforms to the immunity requirements of: IEC 60601-1-2:2007
IEC 61000-4-2 Edition 2.0: 2008-12 Electrostatic Discharge
IEC 61000-4-3 Edition 3.1: 2008-04 Radiated Immunity
IEC 61000-4-8 Edition 1.1:2001-03 Power Frequency Magnetic Fields